

Instructions For Use

© 2025 Beckman Coulter, Inc. All rights reserved.

UniCel DxI Access Wash Buffer II**REF** A16793**FOR PROFESSIONAL USE ONLY****PRINCIPLE****CAUTION**

For U.S.A. only, Federal law restricts this device to sale and distribution by or on the order of a physician, or to a clinical laboratory; and use is restricted to by or on the order of a physician.

INTENDED USE

Access Wash Buffer II is intended for use with the Access Immunoassay Systems and specific Access immunoassay reagents.

REAGENTS**PRODUCT INFORMATION**

- Provided ready to use.
- Stable until the expiration date stated on the label when stored at room temperature (15 to 30°C). An increase in substrate background measurements or increased relative light units for the zero calibrators in "sandwich"-type assays may indicate instability.

R3	Wash Buffer II: 1 x 10 L TRIS buffered saline, surfactant, < 0.1% sodium azide, and < 0.05% Reaction mass of: 5-chloro-2-methyl-4-isothiazolin-3-one and 2-methyl-4-isothiazolin-3-one (3:1).
-----------	---

WARNING AND PRECAUTIONS

- For *in vitro* diagnostic use.
- For hazards presented by the product refer to the following sections: REACTIVE INGREDIENTS and GHS HAZARD CLASSIFICATION.

REACTIVE INGREDIENTS **CAUTION**

Sodium azide preservative may form explosive compounds in metal drain lines. See NIOSH Bulletin: Explosive Azide Hazard (8/16/76). To avoid the possible build-up of azide compounds, flush wastepipes with water after the disposal of undiluted reagent. Sodium azide disposal must be in accordance with appropriate local regulations.

GHS HAZARD CLASSIFICATION

Wash Buffer II

WARNING



H317	May cause an allergic skin reaction.
H412	Harmful to aquatic life with long lasting effects.
P273	Avoid release to the environment.
P280	Wear protective gloves, protective clothing and eye/face protection.
P333+P313	If skin irritation or rash occurs: Get medical advice/attention.
P362+P364	Take off contaminated clothing and wash it before use.

reaction mass of: 5-chloro-2-methyl-4-isothiazolin -3-one [EC# 247-500-7] and 2-methyl-4-isothiazolin-3-one [EC# 220-239-6](3:1) < 0.05%

SDS

Safety Data Sheet is available at beckmancoulter.com/techdocs

TESTING PROCEDURE(S)

DIRECTIONS FOR USE Refer to the appropriate system manuals and/or Help system for detailed instructions.

ADDITIONAL INFORMATION

For a patient/user/third party in the European Union and in countries with identical regulatory regime (Regulation 2017/746/EU on In vitro Diagnostic Medical Devices); if, during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and/or its authorized representative and to your national authority.

Beckman Coulter, the stylized logo, and the Beckman Coulter product and service marks mentioned herein are trademarks or registered trademarks of Beckman Coulter, Inc. in the United States and other countries.

May be covered by one or more pat. -see www.beckmancoulter.com/patents.

REVISION HISTORY

Revision N

New release of IVDR compliant IFU.

Revision P

Added Translations.

Revision R

Added Translations.

Revision T


Added Translations.

SYMBOLS KEY

Glossary of Symbols is available at beckmancoulter.com/techdocs (document number C02724).

EC	REP
----	-----

 Beckman Coulter Ireland Inc., Lismeehan, O'Callaghan's Mills, Co. Clare, Ireland +(353) (0) 65 683 1100

 Beckman Coulter, Inc., 250 S. Kraemer Blvd., Brea, CA 92821 U.S.A.
+(1) 800-854-3633
www.beckmancoulter.com