



# Certificate of Compliance

**Certificate:** 2553553 (LR 34862)

**Master Contract:** 169502

**Project:** 2553553

**Date Issued:** October 3, 2012

**Issued to:** Beckman Coulter Inc.  
250 S. Kraemer Blvd  
Brea, CA 92822  
USA  
Attention: Jeffrey Gonzalez

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Constantine Somow*

**Issued by:** Constantine Somow, CTech

## **PRODUCTS**

**CLASS 8721 05** - LABORATORY EQUIPMENT - Electrical

**CLASS 8721 85** - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Automated Blood Sample Analysis System, Model Power Link 4210, rated 208 V ac, 50/60 Hz, 10 A.

Power Link consists of the following units:

Unit Name	Unit No.
Single Arm Inlet Outlet	9720
De-Capper	9722

## Notes:

- The above models are Equipment Class I, Pollution Degree 2, Installation Category II.
- Mode of operation: Continuous
- Environmental Conditions: Normal: 15 to 40 C, 2000 m max, 15-85% RH



**Certificate:** 2553553 (LR 34862)

**Master Contract:** 169502

**Project:** 2553553

**Date Issued:** October 3, 2012

---

### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-081:04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-081: Particular requirements for automatic and semiautomatic laboratory equipment for analysis and other purposes

CAN/CSA-C22.2 No. 61010-2-101:04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

IEC Standard No. 61010-2-081:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

IEC Standard No. 61010-2-101:2002 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment



## *Supplement to Certificate of Compliance*

**Certificate:** 2553553

**Master Contract:** 169502

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
2553553	October 3, 2012	Automated Blood Sample Analysis System, Model Power Link 4210 (C/US)