BECKMAN COULTER

CERTIFICATE OF ANALYSIS

INSTRUMENT:	AU680 OA47	**PACK LOT NO:**	2550
PRODUCT:	IgG REAGENT	**PRODUCTION DATE**	2023-01-01
CATALOGUE NO :	OSR61172	**EXPIRY DATE:**	2025-01-01

KIT CONCEPT

FILLING VOLUME:
REAGENT 1 : COMPLIES
REAGENT 2 : COMPLIES

PHYSICAL APPEARANCE:
REAGENT 1 : CLEAR COLOURLESS LIQUID
REAGENT 2 : CLEAR LIGHT BROWN AMBER LIQUID

	FRESH
REAGENT BLANK	0.2611

FACTORS / OD'S				
	1.	462.72	/	0.0691
	2.	425.55	/	0.1442
	3.	465.75	/	0.2814
	4.	565.14	/	0.5076
	5.	766.46	/	0.8412
	6.		/	

UNITS: mg/L

CALIBRATOR ***CAT. NO:*** ODR3021 ***PACK LOT NO:*** 1037B

CONTROL SERA RESULTS:

CONTROL SERUM	*L'chek Spinal Fluid Ctrl L1*	*L'chek Spinal Fluid Ctrl L2*	
LOT No.	*56491*	*56492*	
MEAN SD CV%	52.66 0.84 1.59	412.59 2.99 0.72	
TARGET RANGE	*54* *43.2 - 64.8*	*416* *333 - 499*	

QC ANALYST: JASMINE LEMUSIAN
DATE: 02 Feb 2023

APPROVED BY: Riona Ryan
DATE: 02 FEB 2023

61172CSFE.V23

ITA-TDM_COA_V0

Confidential - Company Proprietary