CERTIFICATE OF ANALYSIS

Product: ACCESS 25 (OH) VITAMIN D TOTAL for use on Access 2 Immunoassay System

This certificate is intended for use with COA: ACCESS 25 (OH) VITAMIN D TOTAL for use on Access 2 Immunoassay System

Document No: B30729

Identification

The following information describes this product lot.

Kit Item Number	B24838
Kit Lot Number	439754
Expiration Date	2025-03-31
Date of Manufacture	2024-06-29

Control Values

The following table lists the control lots, stated ranges and results obtained during testing.

Level	QC Lot	Range (ng/mL)	Mean Results (ng/mL)
1	65001	14.4 – 20.3	17.0
2	65002	40.3 – 56.9	47.9
3	65003	98.2 – 124.9	110.9
Level	**Patient Lot**	**Range (ng/mL)**	**Mean Results (ng/mL)**
1	127643	10.6 – 16.3	13.3
2	127644	19.7 – 28.4	23.6
3	127647	38.8-52.5	45.8
4	127648	71.7-97.0	83.4
5	127649	98.2-132.9	119.5

Other Results

The following is a list of additional parameters tested for this product.

Parameter	Specification	Result
Label Information	Correct ID and expiration is listed on the product label	Accept

Approval

Reviewed and Approved By/Date		Title
McGinn, Krystal A	08/02/2024 14:22:15 (UTC)	Quality Control Department

Confidential - Company Restricted

Use or disclosure of information on this page is subject to the restrictions on the title page of this document.