



MIAMI AND KENTUCKY REAGENT MANUFACTURING CERTIFICATE OF ANALYSIS

IDENTIFICATION:	Product Name:	DXH DIFF PACK
	Part Number:	628020
	Lot Number (Finished Product):	3614030
	Lot Number (Sub-Assembly):	110640F, 110793F
	Date of Manufacture:	10-SEP-2024
	Expiration Date:	03-JUN-2025

TEST	SPECIFICATION	RESULTS	Pass/Fail
Functional Testing	NE from parity or ref. $\pm 5\%$	Meets Specification	Pass
Functional Testing	LY from parity or ref. $\pm 5\%$	Meets Specification	Pass
Functional Testing	MO from parity or ref. $\pm 5\%$	Meets Specification	Pass
Functional Testing	EO from parity or ref. $\pm 3\%$	Meets Specification	Pass
Functional Testing	BA from parity or ref. $\pm 3\%$	Meets Specification	Pass
Functional Testing	Mean count - test lot $\geq 7,792$	Meets Specification	Pass
Functional Testing	Mean count - parity lot $\geq 7,792$	Meets Specification	Pass

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

APPROVAL	TITLE	DATE
Signed by: <i>Olivier Ngassam</i>	Quality Assurance Scientist	17-Sep-2024

Signer Name: Olivier Ngassam
 Signing Reason: I am the author of this document
 Signing Time: 17-Sep-2024 | 11:05:42 AM PDT
 E61D88DE66D84474B96B9F78F0A8C6E9