



## Certificate of Analysis – Reagents and Solutions

**Product Name:** ISE BUFFER, 4 x 2000 mL

**Part Number:** AUH1011

**Lot Number:** M411371

**Storage Temperature:** 2 – 25°C

**Date of Manufacture:** 11/21/2024

**Expiration Date:** 3/31/2026

This product was manufactured in accordance with the U.S. FDA Quality System Regulation (GMP)

Test	Specification	Results	Pass/Fail
General Inspection	Components, Labeling, Lot No., and Expiration Date must be legible, complete, and correct	Meets Specification	Pass
Appearance	Free of Visible Particulates	Meets Specification	Pass
<b>Performance Test</b>			
Calcium	Within acceptance range:		Select
Sodium	Within acceptance range: 151-163	158	Pass
Potassium	Within acceptance range: 6.97-7.70	7.38	Pass
Chloride	Within acceptance range: 112-123	118	Pass
Carbon Dioxide	Within acceptance range:		Select

Unused rows may be left blank without use of 'NA'

**Data entered by:** Edgar Rodriguez

Approval	Title	Date (mm/dd/yyyy)
Haley, Catherine L	Senior Quality Control	11/27/2024 14:54:33 (UTC)