



**MIAMI AND KENTUCKY REAGENT MANUFACTURING  
CERTIFICATE OF ANALYSIS**

IDENTIFICATION:	Product Name:	DXH DILUENT
	Part Number:	628017
	Lot Number:	3555420
	Date of Manufacture:	24 Feb 2025
	Expiration Date:	25 Aug 2026

TEST	SPECIFICATION	RESULTS	Pass/Fail
pH	7.00 – 7.20	Meets Specifications	Pass
Osmolality	331 – 341 mOsm/KG	Meets Specifications	Pass
Background Count	≤ 25/0.1 mL	Meets Specifications	Pass
Sodium	222 – 236 mEq/L	Meets Specifications	Pass
Conductivity	19.00 – 20.00 mS/cm @25 C	Meets Specifications	Pass

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

Signed by:	APPROVAL	TITLE	DATE
<i>Pedro Garcia Colon</i>		SR. QUALITY ASSURANCE SCIENTIST	13-Mar-2025

Signer Name: Pedro Garcia Colon  
 Signing Reason: I am the author of this document  
 Signing Time: 13-Mar-2025 | 7:35:00 AM PDT

FDB933D53ED64FE4A2CA0C1E014EFC91