BECKMAN COULTER

CERTIFICATE OF ANALYSIS

INSTRUMENT:	AU680 OA47	**PACK LOT NO:**	2408
PRODUCT:	Urine/CSF Albumin	**PRODUCTION DATE**	2025-03-01
CATALOGUE NO :	B38858	**EXPIRY DATE:**	2026-09-01

KIT CONCEPT

FILLING VOLUME:
REAGENT 1 : COMPLIES
REAGENT 2 : COMPLIES

PHYSICAL APPEARANCE:
REAGENT 1 :Colourless, Clear, free of particulate
REAGENT 2 : Colourless, Visibly free of particulates

REAGENT BLANK ***FRESH*** N/A

FACTORS / OD'S

1.	/	0.0071
2.	/	0.0533
3.	/	0.1526
4.	/	0.3104
5.	/	0.4747
6.	/	

UNITS: mg/L

LINEARITY ***ACHIEVEL*** 475 ***CLAIM:*** 450

CALIBRATOR ***CAT. NO:*** B38859 ***PACK LOT NO:*** 1059A

CONTROL SERA RESULTS:

CONTROL SERUM	*Bio-Rad Liquichek*	*Bio-Rad Liquichek*	
LOT No.	*97611*	*97612*	
MEAN SD CV%	30.19 0.26 0.85	61.35 1.31 2.13	
TARGET RANGE	*29.7* *23.8-35.7*	*60* *48.0-72.0*	

QC ANALYST: K-Gilhool KIERAN GILHOOL
DATE: 09 Apr 2025

APPROVED BY: RIONA RYAN Riona Ryan
DATE: 09 APR 2025.