BECKMAN COULTER

CERTIFICATE OF ANALYSIS

INSTRUMENT:	AU680 OA74	**PACK LOT NO:**	2676
PRODUCT:	TOTAL BILIRUBIN	**PRODUCTION DATE:**	2025-06-01
CATALOGUE NO:	OSR6612	**EXPIRY DATE:**	2026-06-01

KIT CONCEPT

FILLING VOLUME:

Reagent 1 : COMPLIES

Reagent 2 : COMPLIES

PHYSICAL APPEARANCE:

Colour: Clear slightly yellow liquid, no ppt

Blank: Clear colourless liquid, no ppt

	FRESH
REAGENT BLANK	0.001
FACTOR	603

UNITS: µmol/L

CALIBRATION *LOT NO:* 1125 *VALUE:* 53.4

CONTROL SERA RESULTS:

CONTROL SERUM	CSL1 1049	CSL2 1050
FRESH		
MEAN	24.13	111.69
STD	0.05	0.80
CV%	0.22	0.71
TARGET	23	111
RANGE	17-29	82.1-140

QC ANALYST: NIAMH O'RIORDAN **APPROVED BY:** RIONA RYAN Riona Ryan

DATE: 28 Jul. 2025 **DATE:** 28 Jul 2025.

Confidential - Company Proprietary