BECKMAN COULTER

CERTIFICATE OF ANALYSIS

INSTRUMENT:	AU680 OA90	**PACK LOT NO:**	2562
PRODUCT:	IgG REAGENT	**PRODUCTION DATE**	2025-11-01
CATALOGUE NO :	OSR61172	**EXPIRY DATE:**	2027-11-01

KIT CONCEPT

FILLING VOLUME:
REAGENT 1 : COMPLIES
REAGENT 2 : COMPLIES

PHYSICAL APPEARANCE:
REAGENT 1 : CLEAR COLOURLESS LIQUID
REAGENT 2 : CLEAR LIGHT BROWN AMBER LIQUID

REAGENT BLANK ***FRESH*** 0.1972

FACTORS / OD'S

1.	529	/	0.0607
2.	471	/	0.1290
3.	516	/	0.2534
4.	619	/	0.4609
5.	777	/	0.7915
6.		/	

UNITS: mg/L

CALIBRATOR ***CAT. NO:*** ODR3021 ***PACK LOT NO:*** 1040M

CONTROL SERA RESULTS:

CONTROL SERUM	*L'chek Spinal Fluid Ctrl L1*	*L'chek Spinal Fluid Ctrl L2*	
LOT No.	*56581*	*56582*	
MEAN SD CV%	56.83 0.82 1.44	447.38 2.33 0.52	
TARGET RANGE	*52.4* *41.9-62.9*	*442* *354-530*	

QC ANALYST: KO'Qi Q KIARA O'RIORDAN **APPROVED BY:** RIONA Ryan Riona Ryan
DATE: 10.0.2025 **DATE:** 18 NOV 2025

61172CSFE.V26
ITA-TDM_COA_V0

Confidential - Company Proprietary