

**Beckman Coulter**  
 MIAMI AND KENTUCKY REAGENTS MANUFACTURING  
CERTIFICATE OF ANALYSIS

<b>IDENTIFICATION:</b>	Product Name:	BODY FLUID CONTROLS
	Part Number:	628030
	Lot Number (Finished Product):	4515130K
	Lot Number (Sub-assembly):	183234640, 193244640, 203254640
	Date of Manufacture:	28 Apr 2026
	Expiration Date:	11 Aug 2026

TEST	SPECIFICATION	RESULTS	Pass/Fail
Microbial Integrity	No Growth	Meets Specification	Pass
Value Verification	Must meet established limits for appropriate parameters on designated instruments.	Meets Specification	Pass
pH	6.50 - 6.80	Meets Specification	Pass
Osmolality	Level 1: 315-350 Level 2: 317-354 Level 3: 334-370	Meets Specification Meets Specification Meets Specification	Pass Pass Pass
Hemolysis (if applicable)			
Level 2 & Level 3	≤+4 hemolysis	Meets Specification	Pass
Homogeneity (if applicable)	≤ 2 SD limits	Meets Specification	Pass

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

Signed by:	APPROVAL	TITLE	DATE
<i>Pedro Garcia Colon</i>		Manufacturing Quality Supervisor	04-May-2026

Signer Name: Pedro Garcia Colon  
 Signing Reason: I am the author of this document  
 Signing Time: 04-May-2026 | 6:27:10 AM PDT

FDB933D53ED64FE4A2CA0C1E014EFC91