



### Certificate of Analysis – DxC Reagents

**Product Name:** \*RF\*KIT, CHOL FCN 300 CX/LX

**Part Number:** 467825

**Lot Number:** M604612

**Storage Temperature:** 2 - 8°C

**Date of Manufacture:** 4/27/2026

**Expiration Date:** 3/31/2027

This product was manufactured in accordance with the U.S. FDA Quality System Regulation (GMP)

Test	Specification	Results	Pass/Fail
General Inspection	Components, Labeling, Lot No., and Expiration Date must be legible, complete, and correct	Meets Specification	Pass
Clarity or Appearance	Free of Visible Particulates	Meets Specification	Pass
<b>Performance Test</b>			
Sample 1	Within acceptance range: 86.2 to 93.8	88.0	Pass
Sample 2	Within acceptance range: 138.0 to 152.0	141.7	Pass
Sample 3	Within acceptance range: 191.2 to 206.8	194.0	Pass
Sample 4	Within acceptance range: [redacted] to [redacted]	[redacted]	Select

Unused rows may be left blank without use of 'NA'

**Data entered by:** Edgar Rodriguez

Approval	Title	Date (mm/dd/yyyy)
Magtoto, Rowena R	Quality Control Analyst I	05/04/2026 16:54:00 (UTC)