







 Beckman Coulter Ireland, Inc.
 Lismeehan
 O'Callaghan's Mills
 Co. Clare, Ireland
 www.beckmancoulter.com

 Quality Control
 Diagnostics

Commission No. : 180170199 Customer : US Only

	:	800-7212	
	:	7212756	M
	:	2026-04-08	
	:	2027-04-08	


30°C
2°C

Container, Test Strips and Label

Visual control of container, test strips, label and residual moisture

Parameter	: Visual Control of container, test strips and label
Results	: Meets Requirement

<table border="0"> <tr> <td>Parameter</td> <td>: Residual Moisture of Stopper</td> </tr> <tr> <td>Results</td> <td>: 2.94 %</td> </tr> <tr> <td>Upper Limit</td> <td>: 8.00 %</td> </tr> <tr> <td>Lower Limit</td> <td>: - %</td> </tr> </table>	Parameter	: Residual Moisture of Stopper	Results	: 2.94 %	Upper Limit	: 8.00 %	Lower Limit	: - %	<table border="0"> <tr> <td>Parameter</td> <td>: Residual Moisture of Desiccant Pack</td> </tr> <tr> <td>Results</td> <td>: 5.31 %</td> </tr> <tr> <td>Upper Limit</td> <td>: 8.00 %</td> </tr> <tr> <td>Lower Limit</td> <td>: - %</td> </tr> </table>	Parameter	: Residual Moisture of Desiccant Pack	Results	: 5.31 %	Upper Limit	: 8.00 %	Lower Limit	: - %
Parameter	: Residual Moisture of Stopper																
Results	: 2.94 %																
Upper Limit	: 8.00 %																
Lower Limit	: - %																
Parameter	: Residual Moisture of Desiccant Pack																
Results	: 5.31 %																
Upper Limit	: 8.00 %																
Lower Limit	: - %																

Functional Test with Test Urine

Test Strips must meet the criteria as per IRL-UAQC-PCD-0005

Parameter	: Functional Testing with Test Urine
Results	: Meets Requirement

Signature: <u><i>KAROLINA HEROK</i></u>	Date: <u>07 May 2026</u>
Verified By: <u><i>EAMONN LENIHAN</i></u>	Date: <u>07 May 2026</u>
Released by: <u><i>AIMIE TURNER</i></u>	Date: <u>07 MAY 2026</u>